

SCHILLER AG
Altgasse 68, Postfach
CH-6341 Baar, Switzerland
Tel: +41 41 766 42 42
Fax: +41 41 761 08 80

MWST Nr. 212 858
e-mail: sales@schiller.ch
http://www.schiller.ch

SCHILLER
SWITZERLAND

DECLARATION OF CONFORMITY

Manufacturer

SCHILLER AG
Altgasse 68
6341 Baar

Product

Blood Pressure Recorder

Type

BR-102 plus

Serial numbers starting with: 290.00100ff

We, the undersigned, hereby declare that the medical device (class II a) specified above conforms with the Essential Requirements listed in Annex I, of EC Directive 93/42/EEC.

The appendix II of the Directive 93/42/EEC from June 1993 is being applicable.

Applied standards:

IEC 60601-1 EN 1060-1
IEC 60601-1-2 EN 1060-3
IEC 60601-2-30

CE 0123

Baar (Switzerland), 09.03.2005



Markus Bütler
Quality Assurance Manager



THE ART OF DIAGNOSTICS